

检 测 报 告

TEST REPORT

试样名称 一次性使用丁腈检查手套

PRODUCT NAME: Extraction Disposable Nitrile Gloves

规格型号 Acesun-JW0604-S/M/L

TYPE/MODEL: Acesun-JW0604-S/M/L

委托单位 勒拿医疗科技(上海)有限公司

CUSTOMER: LENA Medical Technology (Shanghai) Co., Ltd

检测类别 委托检测

TEST KIND: Entrustment test

苏州美标检测技术有限公司

Suzhou ASTTEK Testing Technology Co., Ltd

In no circumstances shall the company is responsibility extend beyond inspection, testing and reporting upon the samples actually drawn from the bulk and inspected tested and at the sole and exclusive responsibility of the Principal. This test report cannot be reproduced except in full.

检测报告

TEST REPORT

试样名称 PRODUCT NAME	一次性使用丁腈检查手套				
	Extraction Disposable Nitrile Gloves				
型号规格 TYPE/MODEL	Acesun-JW0604-S/M/L				
	Acesun-JW0604-S/M/L				
检测类别 TEST KIND	委托检测	检测编号 CHECK NUMBER	SZ22121902A		
	Entrustment test				
委托单位 CUSTOMER	勒拿医疗科技(上海)有限公司	单位地址 UNIT ADDRESS	上海市普陀区真南路 822 弄 129 号 3 号楼 106-107 室		
	LENA Medical Technology (Shanghai) Co., Ltd		Rm.106-10,Building 3,No. 129,Lane 822, Zhennan Road,Putuo District,Shanghai		
生产单位 MANUFACTUR ER	---	到样日期 SAMPLE DATE	2022.12.12		

取样方式 RECEIPT	寄送	电话号码 TEL.	021-33609386		
	Post				
检测依据 TEST STANDARD	GB 28881-2012 《手部防护 化学品及微生物防护手套》				
	GB 28881-2012 Hand protection-Protective gloves against chemicals and micro-organisms				
检测日期 TEST DATE	2022.12.12-2022.12.19				
检测结论 CONCLUSION	送检产品测试结果符合 GB 28881 相关性能要求。				
	The test results meet the requirements of GB 28881.				
编写 REPORT MAKER	孙磊	审核 CHECK	杨明	批准 APPROVAL	赵一一

In no circumstances shall the company is responsibility extend beyond inspection, testing and reporting upon the samples actually drawn from the bulk and inspected tested and at the sole and exclusive responsibility of the Principal. This test report cannot be reproduced except in full.

检 测 结 果

TEST RESULTS

序号 NO.	检测项目 TESTING ITEMS	标准要求 SPECIFICATIONS	检测结果 TEST RESULTS	判定 DETERMI NATION
1	抗渗透性能 Impermeability	1.样品在要求环境(温度: 22℃, 相对湿度: 65%RH)下放置 24h。 2.选取化学品, 放入渗透测试池, 保持持续搅拌, 检测捕集介质中化学品的含量及时间, 在达到渗透状态后结束试验。 要求: 手套抗渗透等级不低于 2 级。 1. The sample was placed in the required environment (temperature: 22 ℃, relative humidity: 65% RH) for 24h. 2. Select the chemicals, put them into the penetration test tank, keep stirring continuously, detect the content and time of chemicals in the capture medium, and end the test after reaching the penetration state. Requirements: the impermeability grade of gloves shall not be lower than Grade 2.	符合 Accord	P
备注 REMARKS		测试过程均在技术要求规定的环境条件下进行。 The testing process is carried out under the environmental conditions specified by the technical requirements.		
符号含义: P: 检验结果符合要求;F: 检验结果不符合要求;N: 结果不发生;---: 无内容。 "P" represents pass;"F" represents failure;"N" represents not happen;"---" represents not content.				

*****接下页*****

In no circumstances shall the company is responsibility extend beyond inspection, testing and reporting upon the samples actually drawn from the bulk and inspected tested and at the sole and exclusive responsibility of the Principal. This test report cannot be reproduced except in full.

检测结果

TEST RESULTS

检测部位 Test parts	化学物质 Chemical substances	Cas 号 Cas No.	浓度 Concentration %	样品编号 Sample No.		
				1	2	3
手掌 Palm	甲醇 Methanol	67-56-1	99.5	2级 Grade 2	2级 Grade 2	2级 Grade 2
	乙酸乙酯 Ethylacetate	141-78-6	99.0	2级 Grade 2	2级 Grade 2	2级 Grade 2
	正己烷 n-Hexane	110-54-3	97.0	3级 Grade 3	3级 Grade 3	3级 Grade 3
	四氢呋喃 Tetrahydrofuran	141-78-6	99.0	2级 Grade 2	2级 Grade 2	2级 Grade 2
	二氯甲烷 Dichloromethane	75-09-02	99.5	2级 Grade 2	2级 Grade 2	2级 Grade 2
	丙酮 Acetone	67-64-1	99.2	2级 Grade 2	2级 Grade 2	2级 Grade 2
	乙腈 Acetonitrile	75-05-8	99.0	2级 Grade 2	2级 Grade 2	2级 Grade 2
	甲苯 Toluene	108-88-3	99.5	2级 Grade 2	2级 Grade 2	2级 Grade 2
	N,N-二甲基甲酰胺 DMF	68-12-2	99.5	2级 Grade 2	2级 Grade 2	2级 Grade 2
	二硫化碳 Carbondisulfide	75-15-0	99.0	2级 Grade 2	2级 Grade 2	2级 Grade 2
	硫酸 Sulfuricacid,	7664-93-9	96.0	4级 Grade4	4级 Grade4	4级 Grade4
	氢氧化钠 Sodiumhydroxide	1310-73-2	40.0	5级 Grade 5	5级 Grade 5	5级 Grade 5
	三乙胺 Triethylamine	121-44-8	99.5	2级 Grade 2	2级 Grade 2	2级 Grade 2
	三氟乙酸 TFA	76-05-1	99	2级 Grade 2	2级 Grade 2	2级 Grade 2

In no circumstances shall the company is responsibility extend beyond inspection, testing and reporting upon the samples actually drawn from the bulk and inspected tested and at the sole and exclusive responsibility of the Principal. This test report cannot be reproduced except in full.

样 品 图 片
SAMPLE PICTURE



*****接下页*****

声 明

STATEMENT

1. 我司承诺所有的测试数据都保证真实有效,如客户对测试结果有疑义,可以在一周内提出复检。
We promise all the test data are correct and true, the client can request re-test within one week, if doubting the test result.
2. 我司将对客户所有的产品信息进行保密,绝不泄露。
All the information for the samples will be protected by us ,no any will be leaked out.
3. 所测试的样品可以在实验室保存一周,之后将根据实验室的规定进行销毁处理。
The samples can be kept in our lab for one week, later it will be disposed based on the regulation of the Lab.
4. 测试的结果仅作为参照,我们不对除样品外的产品负责。
The test result of the samples are the reference only, we are not responsible for other products.
5. 如有争端,客户可以委托苏州仲裁委员会进行处理。
When disputing, the client can request Suzhou arbitration commission to handle it.

In no circumstances shall the company is responsibility extend beyond inspection, testing and reporting upon the samples actually drawn from the bulk and inspected tested and at the sole and exclusive responsibility of the Principal. This test report cannot be reproduced except in full.